

CALIBRATION CERTIFICATE

上海市计量测试技术研究院
华东国家计量测试中心

校准证书

委托者
Customer

合达电子(苏州)有限公司

委托者地址
Address of customer

苏州工业园区苏虹中路 225 号星虹大厦 2007 室

器具名称
Name of instrument

数码水平仪

制造厂
Manufacturer

JSB TECH PTE LTD

型号/规格
Model/Specification

DWL-600E/0°~90°

器具编号
No. of instrument

SNA09K10033

器具准确度
Instrument accuracy

1

批准人/职务
Approved by / Functions

(机构校准专用章)

核 验 员
Checked by校 准 员
Calibrated by校准日期 2011 年 07 月 22 日
Date for calibrated Year Month Day地址: 上海市张衡路 1500 号(总部)
Address No.1500 Zhangheng Road, Shanghai(headquarters)

电话: 021-38839800

传真: 021-50798390

邮编: 201203

客户咨询电话: 800-820-5172

投诉电话: 021-50798262

国家法定计量检定机构计量授权证书号(中心/院): (国) 法计(2007) 01039号/(2007) 01019号

The number of the Certificate of Metrological Authorization to The Legal Metrological Verification Institution is No. (2007) 01039 / No. (2007) 01019

中国合格评定国家认可委员会实验室认可证书号: No. CNAS L0134

The number of the certificate accredited by CNAS is No. L0134

本次校准所依据的技术规范(代号、名称):

Reference documents for the calibration (code, name)

参照 JJG33-2002 万能角度尺检定规程

本次校准所使用的主要计量标准器具:

Main measurement standards used in this calibration

名称/型号 Name/Model	编号 Number	证书编号/有效期限 Certificate No./Due date	测量范围/准确度等级或 最大允差或不确定度 Measurement range/accuracy class or maximum permissible errors or uncertainty of measurement
角度块/7 块组	95700	2011D31-20-430016/ 2012.01.09	15°10'~90°/1 级

以上计量标准器具的量值溯源至国家基准。

Quantity values of above measurement standards used in this calibration are traced to those of the national primary standards in the P.R. China.

校准地点及环境条件:

Location and environmental condition for the calibration

地点: 院总部机械楼 112 室

Location

温度: 20.2 °C; 湿度: 55.7 %RH; 其它: /

Ambient temperature

Relative humidity

Others

本次校准结果的扩展不确定度:

Expanded uncertainty

 $U=0.05^{\circ}(k=2)$

校准结果/说明:

Results of calibration and additional explanation

见续页

本证书提供的结果仅对本次被校的器具有效。

The data are valid only for the instrument(s).

校准证书续页专用

Continued page of calibration certificate

校准结果/说明 (续页):

Results of calibration and additional explanation (continued page)

受检点	仪器显示值
15.17°	15.10°
30.33°	30.30°
45.50°	45.40°
50.00°	49.95°
60.67°	60.60°
75.83°	75.70°
90.00°	89.95°

校准结果内容结束